

Converus EyeDetect+ V2 Electrodermal Sensor Specifications

- Type of circuit: Constant current
- Measurement units: Ohms
- Excitation: 1.2 microamps
- Recommended electrodes: 11 mm diameter, Ag/AgCl snap electrodes with 0.05-0.075 molar NaCl electrolyte
- Recommended electrode placement: the volar surface of distal or medial phalanges, thenar and hypothenar surfaces of the palm, or two sites at the inner aspect of the foot, over the abductor hallucis muscle adjacent to the sole of the foot and midway between the proximal phalanx of the big toe and a point directly beneath the ankle
- Filtering:
 - Hardware: 30 Hz low-pass
 - Software: 10 Hz low-pass and 0.0159 Hz high-pass (10-sec TC)